



# Calibration certificate

Item No.: 500001/500002

Dimetix AG-Degersheimerstr. 14- CH-9100 Herisau - SWITZERLAND <b>DIMETIX</b>	
<b>Certificate Of Calibration</b>	
Certificate No.: 37202 Manufacturer: Dimetix AG Type: DLS-A 15 Model No.: 500502 Date of Calibration: 07-Januar-2005 Temperature: 20 +/- 4° C	
Description: DLS-A 15 Distance Laser Sensor Serial No.: 44820102 Humidity: 40 +/- 0% RH	
At the completion of the calibration, measured values were in specification stated in the technical reference manual.	
Calibration Equipment Used Model Number: D0002 Model Description: Laser Distance Sensor	
Serial Number: 181144 Calibration Date: 15.Januar 2005	
Herisau, 10-März-2005  _____ Andreas Zuernd Quality Manager	

For a certificate of calibration for a DLS/FLS device, we gauge distances from 0.2 m to 16 m. The reference distances are based on national and international standards.

## Certificate of calibration – for new DLS/FLS measuring devices

When you order new DLS/FLS sensors, you may order a certificate of calibration for each device at the same time.

## Recalibration with calibration certificate – for used DLS/FLS measuring devices

You may return previously used DLS/FLS devices to Dimetix AG for recalibration. These units go through the entire quality process: device cleaning, climatic exposure test cabinet, function test and calibration. Following this, a certificate of calibration will be produced. Any necessary repairs are not included in the price.

## Order information

Item No.	Description	Packaging unit
500001	Calibration certificate for new measuring devices DLS/FLS (must be ordered together with the device)	1 piece
500002	Recalibration with calibration certificate for used measuring devices DLS/FLS	1 piece

07. Jan. 2014 / Subject to change without notice

500001\_002 Calibration certificate\_EN\_V4.odt